



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/823,411	04/12/2004	Scott C. Anderson	003-005-C1	3240

55962 7590 07/09/2008  
SJM/AFD-WILEY  
14901 DEVEAU PLACE  
MINNETONKA, MN 55345-2126

EXAMINER
----------

SHAY, DAVID M

ART UNIT	PAPER NUMBER
----------	--------------

3735

NOTIFICATION DATE	DELIVERY MODE
-------------------	---------------

07/09/2008

ELECTRONIC

**Please find below and/or attached an Office communication concerning this application or proceeding.**

The time period for reply, if any, is set in the attached communication.

Notice of the Office communication was sent electronically on above-indicated "Notification Date" to the following e-mail address(es):

RHEIMBECHER@SJM.COM

<b>Office Action Summary</b>	<b>Application No.</b> 10/823,411	<b>Applicant(s)</b> ANDERSON ET AL.	
	<b>Examiner</b> david shay	<b>Art Unit</b> 3735	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

### Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

### Status

- 1) ☒ Responsive to communication(s) filed on December 10, 2007.
- 2a) ☒ This action is **FINAL**.                      2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

### Disposition of Claims

- 4) ☒ Claim(s) 50-54 and 83-95 is/are pending in the application.
- 4a) Of the above claim(s) 50-54 is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 83-95 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

### Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

### Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All    b) ☐ Some \*    c) ☐ None of:
- ☐ Certified copies of the priority documents have been received.
  - ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  - ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

### Attachment(s)

- |  |   |
|--|---|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892)                     | 4) <input type="checkbox"/> Interview Summary (PTO-413)           |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | Paper No(s)/Mail Date. _____                                      |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08)          | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| Paper No(s)/Mail Date _____  | 6) <input type="checkbox"/> Other: _____                          |

The text of those sections of Title 35, U.S. Code not included in this action can be found in a prior Office action.

Applicant has amended the priority data to reflect the fact that the instant application is a Divisional of U.S. Patent Application No. 09/507,336, rather than a continuation. Since this shift in priority is merely a change in the type of benefit claimed, rather than a new claim for earlier benefit of an earlier application, and further, since the claim for benefit per se was already extant in the instant application, the shift appears to be allowable under 37 C.F.R. 1.78(a)(2)(ii).

The instant application now being filed as a Divisional of US Patent Application Serial No. 09/507,336, a shift of the invention desired to be examined is appropriate. Accordingly, applicant has elected to prosecute group IV, which is now covered by claims 83-95, with claims 50-54 now being withdrawn from consideration.

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 83, 89, 90, and 94 rejected under 35 U.S.C. 102(b) as being clearly anticipated by Choy.

See Figure 1, wherein the distal tip of the device is the outer lip, the laser fiber is the ablating element, the ring of illumination fibers (see Figure 4) constitutes an inner lip, the portion of the device distal of the tip of the optical fiber constitutes a suction well, the annular volume surrounding the fiber constitutes fluid chamber, and the portion of the catheter proximal of the

Art Unit: 3735

fiber is the suction lumen, the suction port being that portion of the catheter coupled to the suction source.

Claims 83, 89, 90, and 94 rejected under 35 U.S.C. 102(e) as being clearly anticipated by Vierra et al.

Claims 84-88, 91-93, and 95 are rejected under 35 U.S.C. 103(a) as being unpatentable over Choy. Choy teaches a device as claimed except the multiple wells, the sensors and the particular ratio of the suction lumen cross section to the vacuum port cross section. It would have been obvious to the artisan of ordinary skill to provide multiple ports for multiplied effect, such as operating on multiple cardiac vessels simultaneously, and to provide a pressure, flow rate, or electrical sensor since these are equivalent to the reservoir 30 of Choy, as they would serve to distinguish when the thrombus is breached and provide no unexpected result, and to provide the particular ratio of the port cross section to the lumen cross section since this is not critical; is well within the skill of one having ordinary skill in the art; and provides no unexpected result, thus producing a device such as claimed.

Claims 84-88, 91-93, and 95 are rejected under 35 U.S.C. 103(a) as being unpatentable over Vierra et al. Vierra et al teaches a device as claimed except the multiple wells, the sensors and the particular ratio of the suction lumen cross section to the vacuum port cross section. It would have been obvious to the artisan of ordinary skill to provide multiple ports for multiplied effect, such as operating on multiple cardiac vessels simultaneously, and to provide a pressure, flow rate, or electrical, sensor since these are equivalent to the reservoir 30 of Choy, as they would serve to distinguish when the thrombus is breached and provide no unexpected result, and to provide the particular ratio of the port cross section to the lumen cross section since this is not

Art Unit: 3735

critical; is well within the skill of one having ordinary skill in the art; and provides no unexpected result, thus producing a device such as claimed.

The nonstatutory double patenting rejection is based on a judicially created doctrine grounded in public policy (a policy reflected in the statute) so as to prevent the unjustified or improper timewise extension of the "right to exclude" granted by a patent and to prevent possible harassment by multiple assignees. See *In re Goodman*, 11 F.3d 1046, 29 USPQ2d 2010 (Fed. Cir. 1993); *In re Longi*, 759 F.2d 887, 225 USPQ 645 (Fed. Cir. 1985); *In re Van Ornum*, 686 F.2d 937, 214 USPQ 761 (CCPA 1982); *In re Vogel*, 422 F.2d 438, 164 USPQ 619 (CCPA 1970); and, *In re Thorington*, 418 F.2d 528, 163 USPQ 644 (CCPA 1969).

A timely filed terminal disclaimer in compliance with 37 CFR 1.321(c) may be used to overcome an actual or provisional rejection based on a nonstatutory double patenting ground provided the conflicting application or patent is shown to be commonly owned with this application. See 37 CFR 1.130(b).

Effective January 1, 1994, a registered attorney or agent of record may sign a terminal disclaimer. A terminal disclaimer signed by the assignee must fully comply with 37 CFR 3.73(b).

Claims 83-95 rejected under the judicially created doctrine of obviousness-type double patenting as being unpatentable over claims 1-25 of U.S. Patent No. 5,830,214. Although the conflicting claims are not identical, they are not patentably distinct from each other because the claims of the patent anticipate the claims of the application. Accordingly, instant application claims are not patentably distinct from the patent application claims. Here, the patent application claims require elements A, B, C, and D while instant application claim 1 only requires elements A, B, and C. Thus it is apparent that the more specific patent application claims encompass the instant application claims. Following the rationale in *In re Goodman* cited in the preceding paragraph, where applicant has once been granted a patent containing a claim for the specific or narrower invention, applicant may not then obtain a second patent with a claim for the generic or broader invention without first submitting an appropriate terminal disclaimer.

Claims 83-95 are provisionally rejected under the judicially created doctrine of obviousness-type double patenting as being unpatentable over claims 1-25 of U.S. Patent

Application No. 11/882,072. Although the conflicting claims are not identical, they are not patentably distinct from each other because the claims of the copending application anticipate the claims of the instant application. Accordingly, instant application claims are not patentably distinct from the copending application claims. Here, the copending application claims require elements A, B, C, and D while instant application claims only requires elements A, B, and C. Thus it is apparent that the more specific copending application claims encompass the instant application claims. Following the rationale in *In re Goodman* cited in the preceding paragraph, where applicant has once been granted a patent containing a claim for the specific or narrower invention, applicant may not then obtain a second patent with a claim for the generic or broader invention without first submitting an appropriate terminal disclaimer.

This is a provisional obviousness-type double patenting rejection because the conflicting claims have not in fact been patented.

Applicant's arguments with respect to claims 83-95 have been considered but are moot in view of the new ground(s) of rejection.

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37

CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to david shay whose telephone number is (571) 272-4773. The examiner can normally be reached on Tuesday through Friday from 6:30 a.m. to 5:00 p.m.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Charles Marmor, II, can be reached on Monday, Tuesday, Wednesday, Thursday, and Friday. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

/david shay/

Primary Examiner, Art Unit 3735